

## Determination of pressure in aqueo-carbonic fluid inclusions at high temperatures from measured Raman frequency shifts of CO<sub>2</sub>

XUEYIN YUAN<sup>1,2</sup>, ROBERT A. MAYANOVIC<sup>3</sup>, HAIFEI ZHENG<sup>1,\*</sup>, AND QIANG SUN<sup>1</sup>

<sup>1</sup>Key Laboratory of Orogenic Belts and Crustal Evolution, Ministry of Education, Peking University, Beijing 100871, China

<sup>2</sup>MLR Key Laboratory of Metallogeny and Mineral Assessment, Institute of Mineral Resources, Chinese Academy of Geological Sciences, Beijing 100037, China

<sup>3</sup>Department of Physics, Astronomy and Material Science, Missouri State University, Springfield, Missouri 65897, U.S.A.

### ABSTRACT

Due to the presence of additional volatiles and/or electrolytes in CO<sub>2</sub>-H<sub>2</sub>O fluids, the total pressure of many natural aqueo-carbonic fluid inclusions at high temperatures as determined using microthermometry is usually made with considerable uncertainty. In this paper, we present the results of our high *P-T* in situ Raman scattering study of high-density aqueo-carbonic fluids, with and without a small amount of CH<sub>4</sub> and NaCl, whose objective is to derive a new method for pressure determination in aqueo-carbonic fluid inclusions at high temperatures. The measurement of the Fermi dyad bands at temperatures up to 400 °C and pressures up to 1200 MPa is described. The manner in which the frequency shifts and intensity of Raman bands are governed by pressure, temperature, presence of CH<sub>4</sub> in carbonic and NaCl in aqueous fluids is discussed. From the monotonic dependence of the frequency shifts of the lower Fermi dyad band  $\nu_-$  and the Fermi resonant splitting  $D$  ( $D = \nu_+ - \nu_-$ ) with pressure and temperature, the pressure (in MPa) in aqueo-carbonic fluid inclusions at elevated temperatures can be determined directly by using the following two polynomial equations:

$$P \text{ (MPa)} = -16 + 1.232 \times T - 53.72 \times (\Delta\nu_-) - 1.83 \times 10^{-3} \times T^2 + 24.46 \times (\Delta\nu_-)^2 - 0.292 \times T \times (\Delta\nu_-),$$
$$P \text{ (MPa)} = -26 + 1.501 \times T + 193.24 \times (\Delta D) - 1.61 \times 10^{-3} \times T^2 + 5.436 \times (\Delta D)^2 + 0.158 \times T \times (\Delta D),$$

where  $T$  is in °C,  $\Delta\nu_-$  and  $\Delta D$  represent frequency shifts (in cm<sup>-1</sup>) of the lower band and the resonant splitting relative to the reference values measured at 23 °C and 6 MPa, respectively. Based on the attainable accuracy of the fitted peak positions and the results from fitting of Raman frequency shifts' dependence with pressure and temperature, the uncertainty in pressure determination is about 50 MPa for pressures determined from  $\nu_-$  and 40 MPa from that determined from  $D$ .

**Keywords:** CO<sub>2</sub>, Raman spectroscopy, pressure determination, HDAC, fluid inclusion